



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

DORSEY METROLOGY CALIBRATION LABORATORY  
53 Oakley Street  
Poughkeepsie, NY 12601  
Andrew Totten Phone: 845 454 3111

CALIBRATION

Valid To: November 30, 2024

Certificate Number: 2981.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory at the location listed above as well as the satellite laboratory location listed below to perform the following calibrations<sup>1</sup>:

I. Dimensional

Parameter/Equipment	Range	CMC <sup>2,4</sup> (±)	Comments
Dial Indicator – High Amplification (High Resolution) Course Graduation (Low Resolution)	Up to 0.050 in Up to 1 in	23 µin + 0.6R 110 µin + 0.6R	MET-7.2.1, Mitutoyo calibration tester
Digital Indicator	Up to 1 in	54 µin	MET-7.2.2, gage blocks
ID/OD Gage and Set Master	Up to 48 in OD Up to 49 in ID Length	13L µin + 0.6R 13L µin + 0.6R	MET-7.2.3, gage blocks

Parameter/Equipment	Range	CMC <sup>2, 4</sup> (±)	Comments
Thickness Gage and Set Master	Up to 6 in	150 μin + 0.6R	MET-7.2.4, gage blocks
Depth/Height Gage and Set Master	Up to 12 in	190 μin + 0.6R	MET-7.2.5, gage blocks
Bore Gage –  0.0001 in (0.0025 mm) Graduation  0.00025 in (0.0064 mm) Graduation  0.0005 in (0.0127 mm) Graduation  0.001 in Graduation	(1 to 36) in  (1 to 36) in  (1 to 36) in  (1 to 36) in	75 μin  160 μin  290 μin  580 μin	MET-7.2.6, bore gage calibrator

SATELLITE LOCATION

DORSEY METROLOGY CALIBRATION LABORATORY  
 25 Oakley Street  
 Poughkeepsie, NY 12601

I. Dimensional

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Optical Comparator <sup>3</sup> –			MET-7.2.7
Squareness	Up to 9 in	130 μin	Perpendicular master and dial test indicator
Magnification	Up to 24 in (X and Y Axis)	80 μin	Magnification glass master
X,Y Linear Measurement	Up to 24 in (X and Y Axis)	150 + 86 x L / 24	Projection glass master

<sup>1</sup> This laboratory offers commercial calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer’s device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for these calibrations. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer’s site being larger than the CMC.

<sup>4</sup> In the statement of CMC,  $L$  is the numerical value of the nominal length of the device measured in inches, and  $R$  is the resolution of the unit under test.



# Accredited Laboratory

A2LA has accredited

## DORSEY METROLOGY CALIBRATION LABORATORY

*Poughkeepsie, NY*

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 10<sup>th</sup> day of November 2022.

A blue ink signature of a person, written in a cursive style, positioned above a horizontal line.

Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 2981.01  
Valid to November 30, 2024  
Revised August 16, 2023

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*